

UV-VIS Double Beam Spectrophotometer

Make of UV-VIS Double Beam Spectrophotometer:

Thermo Fisher Scientific

Model: Spectrascan UV-2600

Type of analysis possible:

- Qualitative Analysis
- Quantitative Analysis



- Double beam UV-VIS Spectrophotometer, data, method and info banks with high speed scanning.
- Kinetics,
- Quantitative analysis,
- Time scan
- Multicomponent analysis.
- Window based software.

Specification:

- Wavelength range: 200-1100 nm (Standard),
190-1100 nm (Optional)
- Monochromator: 1200 lines/nm.

Contact us:

Mobile No.

Email ID: Pankaj.Mishra@icar.gov.in

- | | |
|--|--|
| <ul style="list-style-type: none">• Spectral bandwidth: 1.6 nm/1 nm.• Photometric accuracy: ± 0.005 abs at 1 abs.• Detector: Silicon Photodiode.• Auto zero function. | |
| <p><i>User Instructions :</i></p> <ul style="list-style-type: none">• Qualitative and quantitative data can be provided,• All sample should be homogenous.• Samples should not be toxic or hazardous. | |